

Goldline TEF Division

TEF25 USB Preamplifier – Low Noise Capability Verification



TEST

Goldline

Device

TEF25 USB Preamp

Report Date:

Manufacturer:

October 4, 2005 Test Date: September 27,

2005

PREPARED FOR

Greg Miller Goldline, TEF Division

P.O. Box 500

West Redding, CT 06896 Voice (203) 938-2588 (203)-938-8740 Fax Email sales@gold-line.com **RESULT SUMMARY**

NC = 0

PREPARED BY

David M. Berg

Orfield Laboratories, Inc. 2709 East 25th Street Minneapolis MN 55406 Voice (612) 721-2455 Fax (612) 721-2457 e-mail dave@orfieldlabs.com

The information included in the following report are the results of low noise testing performed on one TEF25 USB Preamp.

Prepared by:

Reviewed by:

David M. Berg **Laboratory Manager**

Elliott B. Dick **Quality Manager**













TEST RESULTS

Measurement Equipment:

Manufacturer: Goldline, TEF Division Specimen: TEF25 USB Preamp

Serial #: **05H1604**

Date Tested: September 27, 2005

Calibrator: Brüel & Kjær type 4230 #1379712*
Microphone: Brüel & Kjær type 4179 #1832800*
Preamplifier: Brüel & Kjær type 2660 #1885025
Power Supply: Brüel & Kjær type WB 1057

*NIST Traceable Calibration on file

Environment:

Anechoic Chamber at Orfield Laboratories.

Conclusions:

When combined with an appropriate very low noise microphone assembly and preamplifier, such as the B&K type 4179 / 2660 combination, the TEF25 USB Preamp was found to be capable of measuring sound pressure levels well below the threshold of human hearing (i.e. NC=0). As is most frequently the case the measurement noise floor capability of the system is dictated by the self-noise of the microphone and preamplifier. Typical high quality (ANSI type 1) ½" and 1" microphones are limited to a noise floor of around 10 – 15 dBA, quiet enough for most any architectural acoustic measurement purpose.




Figure 1 - Noise Floor at Anechoic Chamber w/ TEF25 B&K 4179/2660